

BID EVALUATION SHEET/TECHNICAL SHEET

NAME OF THE EQUIPMENT: ECG MACHINES

PART- I

KNOCK DOWN CRITERIA - (COMMERCIAL EVALUATION)

(To be evaluated by Purchase Department)

(All evaluation parameters defined below are mandatory for compliance)

Sr. No.	Evaluation Parameters	M/s Radiant Medical Pvt. Ltd.	M/s Bio Tech Services	M/s KASBN International
1	Complete Package/Tender	YES	YES	YES
2	Original Receipt of Tender	YES	YES	YES
3	Affidavit from Bidder	YES	YES	YES
4	Bid Security	Photocopy attached	Photocopy attached	Photocopy attached
5	Bid Validity	YES	YES	YES
6	Delivery Period	YES	YES	YES
	Remarks:	Eligible for further part II	Eligible for further part II	Eligible for further part II

h m c

g


PART- II**KNOCK DOWN CRITERIA - (VENDOR EVALUATION)****(To be evaluated by Technical Evaluation Committee)****(All evaluation parameters defined below are mandatory for compliance.)**

Sr. No.	Evaluation Parameters	M/s Radiant Medical Pvt. Ltd.	M/s Bio Tech Services	M/s KASBN International
1	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes
2	Technical & Engineering capability(As defined for the specific tender in specifications)	Yes	Yes	Yes
3	Certificate from the Manufacturer about the after sales services through agent or itself (In case specifically demanded in the specifications)	Yes	Yes	Yes
4	Vendor Past performance (In case of unsatisfactory performance, details must be mentioned)	Yes	Yes	Yes
5	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes
	Remarks:	Eligible for further part III	Eligible for further part III	Eligible for further part III


h
Cm G
Q
g

PART – III
KNOCK DOWN CRITERIA - PRODUCT EVALUATION
 (All evaluation parameters defined below are mandatory for compliance.)


Item Sr. No.	SPECIFICATION COMPLIANCE /EVALUATION PARAMETERS			
	COMPANY	M/s Radiant Medical Pvt. Ltd.	M/s Bio Tech Services	M/s KASBN International
1	Manufacturer	Baxter	Innomed	M4Medical sp
	Model	ELI-230	Heart Screen 112C-1	M-Trace
Country of Manufacturer		USA	Hungary	Poland
Country of Origin of Product/Model Number		USA	Hungary	Poland
Compliance with defined quality standards		Yes	Yes	Yes
Specification Compliance features wise:		Remarks	Remarks	Remarks
Specifications:				
Technical Eligibility of Product:		Eligible	Eligible	Eligible
Technical Eligibility of Firm:		Eligible	Eligible	Eligible
BID STATUS:		Responsive	Responsive	Responsive




Mr. Haseeb Aslam
 Bio Medical Engineer
 Allied Hospital, Faisalabad.




Electronic Equipment Care Taker
 F.I.C, Faisalabad.




Pharmacist
 F.I.C. Faisalabad



D.M.S (Purchase)
 F.I.C, Faisalabad.



A.M.S (Purchase)
 F.I.C, Faisalabad.



Dr. Shahid Abbas,
 Prof. of Cardiology
 F.I.C, Faisalabad.

BID EVALUATION SHEET/TECHNICAL SHEET

NAME OF THE EQUIPMENT: DEFIBRILLATOR

PART-I

KNOCK DOWN CRITERIA - (COMMERCIAL EVALUATION)

(To be evaluated by Purchase Department)

(All evaluation parameters defined below are mandatory for compliance)

Sr. No.	Evaluation Parameters	M/s Medical Equipment & Systems	M/s Bio Tech Services	M/s Radiant Medical Pvt. Ltd.
1	Complete Package/Tender	YES	YES	YES
2	Original Receipt of Tender	YES	YES	YES
3	Affidavit from Bidder	YES	YES	YES
4	Bid Security	Photocopy attached	Photocopy attached	Photocopy attached
5	Bid Validity	YES	YES	YES
6	Delivery Period	YES	YES	YES
	Remarks:	Eligible for further part II	Eligible for further part II	Eligible for further part II

h

Q

CM C
9

PART- II
KNOCK DOWN CRITERIA - (VENDOR EVALUATION)
 (To be evaluated by Technical Evaluation Committee)
 (All evaluation parameters defined below are mandatory for compliance.)

Sr. No.	Evaluation Parameters	M/s Medical Equipment & Systems	M/s Bio Tech Services	M/s Radiant Medical Pvt. Ltd.
1	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes
2	Technical & Engineering capability(As defined for the specific tender in specifications)	Yes	Yes	Yes
3	Certificate from the Manufacturer about the after sales services through agent or itself (In case specifically demanded in the specifications)	Yes	Yes	Yes
4	Vendor Past performance (In case of unsatisfactory performance, details must be mentioned)	Yes	Yes	Yes
5	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes
	Remarks:	Eligible for further part III	Eligible for further part III	Eligible for further part III

h


Er G Q

PART – III
KNOCK DOWN CRITERIA - PRODUCT EVALUATION
 (All evaluation parameters defined below are mandatory for compliance.)

Item Sr. No.	SPECIFICATION COMPLIANCE /EVALUATION PARAMETERS			
	COMPANY	M/s Medical Equipment & Systems	M/s Bio Tech Services	M/s Radiant Medical Pvt. Ltd.
1	Manufacturer	Zoll Medical	Innomed	OSATU, S.COOP
	Model	M2 with pacing	Cardio-Aid 360B	Reanibex-700
Country of Manufacturer		USA	Hungary	Spain
Country of Origin of Product/Model Number		China	Hungary	Spain
Compliance with defined quality standards		No FDA 510K	Yes	Yes
Specification Compliance features wise:		Remarks	Remarks	Remarks
Specifications:				
Technical Eligibility of Product:		Not-Eligible	Eligible	Eligible
Technical Eligibility of Firm:		Eligible	Eligible	Eligible
BID STATUS:		Non-Responsive	Responsive	Responsive



Mr. Haseeb Aslam
 Bio Medical Engineer
 Allied Hospital, Faisalabad.


Electronic Equipment Care Taker
 F.I.C, Faisalabad.


Pharmacist
 F.I.C. Faisalabad


D.M.S (Purchase)
 F.I.C, Faisalabad.


A.M.S (Purchase)
 F.I.C, Faisalabad.


Dr. Shahid Abbas,
 Prof. of Cardiology
 F.I.C, Faisalabad.

BID EVALUATION SHEET/TECHNICAL SHEET
NAME OF THE EQUIPMENT: ECHOCARDIOGRAPHY MACHINE

PART-I

KNOCK DOWN CRITERIA - (COMMERCIAL EVALUATION)

(To be evaluated by Purchase Department)

(All evaluation parameters defined below are mandatory for compliance)

Sr. No.	Evaluation Parameters	M/s Vertex Medical	M/s Bio Tech Services	M/s Eastern Medical Technology Services
1	Complete Package/Tender	YES	YES	YES
2	Original Receipt of Tender	YES	YES	YES
3	Affidavit from Bidder	YES	YES	YES
4	Bid Security	Photocopy attached	Photocopy attached	Photocopy attached
5	Bid Validity	YES	YES	YES
6	Delivery Period	YES	YES	YES
	Remarks:	Eligible for further part II	Eligible for further part II	Eligible for further part II

Q h on C

g


PART- II**KNOCK DOWN CRITERIA - (VENDOR EVALUATION)****(To be evaluated by Technical Evaluation Committee)****(All evaluation parameters defined below are mandatory for compliance.)**


Sr. No.	Evaluation Parameters	M/s Vertex Medical	M/s Bio Tech Services	M/s Eastern Medical Technology
1	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes
2	Technical & Engineering capability(As defined for the specific tender in specifications)	Yes	Yes	Yes
3	Certificate from the Manufacturer about the after-sales services through agent or itself (In case specifically demanded in the specifications)	Yes	Yes	Yes
4	Vendor Past performance (In case of unsatisfactory performance, details must be mentioned)	Yes	Yes	Yes
5	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes
	Remarks:	Eligible for further part III	Eligible for further part III	Eligible for further part III

Q h Am C
g


PART – III
KNOCK DOWN CRITERIA - PRODUCT EVALUATION
 (All evaluation parameters defined below are mandatory for compliance.)

Item Sr. No.	SPECIFICATION COMPLIANCE /EVALUATION PARAMETERS			
	COMPANY	M/s Vertex Medical	M/s Bio Tech Services	M/s Eastern Medical Technology Services
1	Manufacturer	Siemens	Esaote,	GE Healthcare
	Model	Acuson NX3 Elite	MyLab X75	Vivid T9
Country of Manufacturer		Germany	Italy	USA
Country of Origin of Product/Model Number		Slovakia	Italy	Not mentioned
Compliance with defined quality standards		Yes	Yes	Yes
Specification Compliance features wise:		Remarks	Remarks	Remarks
Specifications:		—	End User Reservations (Dual Screen not available, software not user friendly, Image resolution not upto the mark)	—
Technical Eligibility of Product:		Eligible	Not Eligible	Eligible
Technical Eligibility of Firm:		Eligible	Eligible	Eligible
BID STATUS:		Responsive	Non Responsive	Responsive



 Mr. Haseeb Aslam
 Bio Medical Engineer
 Allied Hospital, Faisalabad.


 Electronic Equipment Care
 Taker
 F.I.C, Faisalabad.


 Pharmacist
 F.I.C, Faisalabad


 D.M.S (Purchase)
 F.I.C, Faisalabad.


 A.M.S (Purchase)
 F.I.C, Faisalabad.


 Dr. Shahid Abbas,
 Prof. of Cardiology
 F.I.C, Faisalabad.

BID EVALUATION SHEET/TECHNICAL SHEET

NAME OF THE EQUIPMENT: COLOR DOPPLER

PART - I

KNOCK DOWN CRITERIA - (COMMERCIAL EVALUATION)

(To be evaluated by Purchase Department)

(All evaluation parameters defined below are mandatory for compliance)

Sr. No.	Evaluation Parameters	M/s Medical Equipment & Systems	M/s Bio Tech Services
1	Complete Package/Tender	YES	YES
2	Original Receipt of Tender	YES	YES
3	Affidavit from Bidder	YES	YES
4	Bid Security	May be in Financial	Photocopy attached
5	Bid Validity	YES	YES
6	Delivery Period	YES	YES
	Remarks:	Eligible for further part II	Eligible for further part II

Q

Q

Q

9

PART- II**KNOCK DOWN CRITERIA - (VENDOR EVALUATION)****(To be evaluated by Technical Evaluation Committee)****(All evaluation parameters defined below are mandatory for compliance.)**

Sr. No.	Evaluation Parameters	M/s Medical Equipment & Systems	M/s Bio Tech Services
1	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes
2	Technical & Engineering capability(As defined for the specific tender in specifications)	Yes	Yes
3	Certificate from the Manufacturer about the after sales services through agent or itself (In case specifically demanded in the specifications)	Yes	Yes
4	Vendor Past performance (In case of unsatisfactory performance, details must be mentioned)	Yes	Yes
5	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes
	Remarks:	Eligible for further part III	Eligible for further part III

Q


M


G


9


PART – III
KNOCK DOWN CRITERIA - PRODUCT EVALUATION
 (All evaluation parameters defined below are mandatory for compliance.)


Item Sr. No.	SPECIFICATION COMPLIANCE /EVALUATION PARAMETERS		
	COMPANY	M/s Medical Equipment & Systems	M/s Bio Tech Services
1	Manufacturer	Canon	Esaote,
	Model	Aplio me (CUS-AME00)	MyLab X8 exp
Country of Manufacturer		Japan	Italy
Country of Origin of Product/Model Number		Japan	Italy
Compliance with defined quality standards		Yes	Yes
Specification Compliance features wise:		Remarks	Remarks
Specifications:			End user reservations (Abdomen scan not satisfactory)
Technical Eligibility of Product:		Eligible	Not Eligible
Technical Eligibility of Firm:		Eligible	Eligible
BID STATUS:		Responsive	Non Responsive



 Mr. Haseeb Aslam
 Bio Medical Engineer
 Allied Hospital, Faisalabad.


 Electronic Equipment Care
 Taker
 F.I.C, Faisalabad.


 Pharmacist
 F.I.C. Faisalabad


 D.M.S (Purchase)
 F.I.C, Faisalabad.


 A.M.S (Purchase)
 F.I.C, Faisalabad.


 Dr. Shahid Abbas,
 Prof. of Cardiology
 F.I.C, Faisalabad.

BID EVALUATION SHEET/TECHNICAL SHEET


NAME OF THE EQUIPMENT: CRASH CART TROLLEY

PART- I

KNOCK DOWN CRITERIA - (COMMERCIAL EVALUATION)

(To be evaluated by Purchase Department)


(All evaluation parameters defined below are mandatory for compliance)

Sr. No.	Evaluation Parameters	M/s Novatek Pakistan
1	Complete Package/Tender	YES
2	Original Receipt of Tender	YES
3	Affidavit from Bidder	YES
4	Bid Security	Photocopy attached
5	Bid Validity	YES
6	Delivery Period	YES
	Remarks: 	Eligible for further part II



PART- II**KNOCK DOWN CRITERIA - (VENDOR EVALUATION)****(To be evaluated by Technical Evaluation Committee)****(All evaluation parameters defined below are mandatory for compliance.)**

Sr. No.	Evaluation Parameters	M/s Novatek Pakistan
1	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
2	Technical & Engineering capability(As defined for the specific tender in specifications)	Yes
3	Certificate from the Manufacturer about the after sales services through agent or itself (In case specifically demanded in the specifications)	Yes
4	Vendor Past performance (In case of unsatisfactory performance, details must be mentioned)	Yes
5	Availability of relevant Tools and Testing / Calibration Equipment	Yes
6	Compliance of Warranty as per tender	Yes
	Remarks:	Eligible for further part III



PART – III
KNOCK DOWN CRITERIA - PRODUCT EVALUATION
 (All evaluation parameters defined below are mandatory for compliance.)

Item Sr. No.	SPECIFICATION COMPLIANCE /EVALUATION PARAMETERS	
	COMPANY	M/s Novatek Pakistan
1	Manufacturer	Beijing Jingdong Tech (Laoting co. Ltd.)
	Model	N6
Country of Manufacturer		China
Country of Origin of Product/Model Number		China
Compliance with defined quality standards		Yes
Specification Compliance features wise:		Remarks
Specifications:		
Technical Eligibility of Product:		Eligible
Technical Eligibility of Firm:		Eligible
BID STATUS:		Responsive



Mr. Haseeb Aslam
 Bio Medical Engineer
 Allied Hospital, Faisalabad.




Electronic Equipment Care Taker
 F.I.C, Faisalabad.



Pharmacist
 F.I.C. Faisalabad



D.M.S (Purchase)
 F.I.C, Faisalabad.



A.M.S (Purchase)
 F.I.C, Faisalabad.



Dr. Shahid Abbas,
 Prof. of Cardiology
 F.I.C, Faisalabad.

BID EVALUATION SHEET/TECHNICAL SHEET

NAME OF THE EQUIPMENT: CARDIAC MONITORS

PART - I

KNOCK DOWN CRITERIA - (COMMERCIAL EVALUATION)

(To be evaluated by Purchase Department)

(All evaluation parameters defined below are mandatory for compliance)

Sr. No.	Evaluation Parameters	M/s Friends Traders	M/s Bio Tech Services	M/s Medical Equipment & Systems	M/s Radiant Medical Pvt. Ltd.	M/s Clinical Life
1	Complete Package/Tender	YES	YES	YES	YES	YES
2	Original Receipt of Tender	YES	YES	YES	YES	YES
3	Affidavit from Bidder	YES	YES	YES	YES	YES
4	Bid Security	Photocopy attached	Photocopy attached	Photocopy attached	Photocopy attached	Photocopy attached
5	Bid Validity	YES	YES	YES	YES	YES
6	Delivery Period	YES	YES	YES	YES	YES
	Remarks:	Eligible for further part II	Eligible for further part II	Eligible for further part II	Eligible for further part II	Eligible for further part II

Handwritten signature and initials in blue ink.

PART - II
KNOCK DOWN CRITERIA - (VENDOR EVALUATION)
 (To be evaluated by Technical Evaluation Committee)
 (All evaluation parameters defined below are mandatory for compliance.)

Sr. No.	Evaluation Parameters	M/s Friends Traders	M/s Bio Tech Services	M/s Medical Equipment & Systems	M/s Radiant Medical Pvt. Ltd.	M/s Clinical Life
1	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	No legal attested authorization certificate	Yes	Yes	No legal attested authorization certificate
2	Technical & Engineering capability(As defined for the specific tender in specifications)	Yes	Yes	Yes	Yes	Yes
3	Certificate from the Manufacturer about the after sales services through agent or itself (In case specifically demanded in the specifications)	Yes	Yes	Yes	Yes	Yes
4	Vendor Past performance (In case of unsatisfactory performance, details must be mentioned)	Yes	Yes	Yes	Yes	Yes
5	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes	Yes	Yes
	Remarks:	Eligible for further part III	Not Eligible for further part III	Eligible for further part III	Eligible for further part III	Not Eligible for further part III

9 ✓ 0 20 C

PART - III

KNOCK DOWN CRITERIA - PRODUCT EVALUATION

(All evaluation parameters defined below are mandatory for compliance.)

SPECIFICATION COMPLIANCE /EVALUATION PARAMETERS							
Item Sr. No.	COMPANY	M/s Friends Traders	M/s Bio Tech Services	M/s Medical Equipment & Systems	M/s Radiant Medical Pvt. Ltd.	M/s Clinical Life	
1	Manufacturer	Mindray	Edan Instruments	Space Lab Healthcare	Biolight	Edan Instruments	
	Model	UMEC 15	iM-80	XPREZZON 91393	Q7	iM-80	
Country of Manufacturer		China	China	USA	China	China	
Country of Origin of Product/Model Number		China	China	USA	China	China	
Compliance with defined quality standards		Yes	Yes	Yes	No	Yes	
Specification Compliance features wise:		Remarks	Remarks	Remarks	Remarks	Remarks	
Specifications:					Arrhythmia analysis are not for paediatric patients as mentioned on FDA 510K certificate		
Technical Eligibility of Product:		Eligible	Eligible	Eligible	Not Eligible	Eligible	
Technical Eligibility of Firm:		Eligible	Not Eligible	Eligible	Eligible	Not Eligible	
BID STATUS:		Responsive	Non Responsive	Responsive	Non Responsive	Non Responsive	

Mr. Haseeb Aslam
Bio Medical Engineer
Allied Hospital, Faisalabad.

Electronic Equipment Care
Taker
F.I.C, Faisalabad.

Pharmacist
F.I.C. Faisalabad

D.M.S (Purchase)
F.I.C, Faisalabad.

A.M.S (Purchase)
F.I.C, Faisalabad.

Dr. Shahid Abbas,
Prof. of Cardiology
F.I.C, Faisalabad.