

**MINUTES OF GRIEVANCE COMMITTEE MEETING
FAISALABAD INSTITUTE OF CARDIOLOGY, FAISALABAD**

A meeting of Grievance Committee for Personal Protection Equipments under Emergency Procurement was held on 18.04.2020 at 10:00AM in the Office of the Medical Superintendent, Faisalabad Institute of Cardiology, Faisalabad.

1. M/s Ali Medical Store

Item #	Item Name	TAC Observations	GC Observation	Remarks
1	Long Sleeve Water resistance Gown PPE's Kits	No Water Resistance	End user Reservation	Rejected
4	Mask Grade N95/N99/ equivalent	Fitting not proper	End user Reservation	Rejected

2. M/s Husnain Pharma

Item #	Item Name	TAC Observations	GC Observation	Remarks
1	Long Sleeve Water resistance Gown PPE's Kits	No Water Resistance	End user Reservation	Rejected
4	Mask Grade N95/N99	Fitting not proper	End user Reservation	Rejected

3. M/s Usman & Company

Item #	Item Name	TAC Observations	GC Observation	Remarks
4	Mask Grade N95/N99	Fitting not proper	End user Reservation	Rejected

M.A.P.

[Signature]

[Signature]

[Signature]

4. M/s Friends Enterprises

Item #	Item Name	TAC Observations	GC Observation	Remarks
1	Long Sleeve Water resistance Gown PPE's Kits	No water resistant	End user Reservation	Rejected
8	Ordinary/Local Surgical Mask (with rubber band)	Poor Quality	End user Reservation	Rejected
9	Disposable pillow cover water resistant	No water resistant	End user Reservation	Rejected
10	Disposable bed sheet water resistant	No water resistant	End user Reservation	Rejected

5. M/s Ibrahim Enterprises

Item #	Item Name	TAC Observations	GC Observation	Remarks
1	Long Sleeve Water resistance Gown PPE's Kits	No water resistant	End user Reservation	Rejected
3	Eye Protection Goggles	Poor Quality	End user Reservation	Rejected
4	Mask Grade N95/N99	Poor Quality Not fit properly	End user Reservation	Rejected
9	Disposable pillow cover water resistant	Poor Quality No water resistant	End user Reservation	Rejected
10	Disposable bed sheet water resistant	Poor Quality No water resistant	End user Reservation	Rejected


6. M/s Innovative Energy Solution


Item #	Item Name	TAC Observations	GC Observation	Remarks
3	Eye Protection Goggles	Sample not provided	Sample not provided	Rejected
4	Mask Grade N95/N99	Sample not provided	Sample not provided	Rejected
7	Face Shield	Sample not provided	Sample not provided	Rejected


Handwritten signature and initials in blue ink.

7. M/s Inova International


Item #	Item Name	TAC Observations	GC Observation	Remarks
	Bid Rejected	CNIC Not Attached Bank Statement not attached Affidavit on judicial paper not attached Bid form (sign & Stamp) not attached	The firm provided the documents but this tender is Emergency Procurement due to spread of Pandemic COVID-19. The relevant documents are mandatory at the time of submission of tender. At this stage it requires more time to evaluate the documents as well as technical evaluation.	The GC unanimously rejected the firm's bid.


Ms. Ayesha Sultana,
Pharmacist,
FIG, Faisalabad.


Mrs. Iqra Munir,
Pharmacist,
FIG, Faisalabad.


Mr. Zaman Mhikhar,
BME Engineer,
FIG, Faisalabad.

Dr. Shahbaz Ahmed,
Associate Prof. of Cardiac Surgery,
FIG, Faisalabad.


Dr. Shahid Abbas,
Associate Prof. of Cardiology,
FIG, Faisalabad.


Dr. Zafar Abbas Khan,
Medical Superintendent,
FIG, Faisalabad.