



SHOE COVER MACHINE
FAISALABAD INSTITUTE OF CARDIOLOGY, FAISALABAD.


Specification	Qty
<p>Shoe Cover Machine</p> <ul style="list-style-type: none">• Operation should be automatic using thermo shrinkable technology or auto cover technique, only covers bottom of the shoes• Capable to handle paper/shoe cover with thickness of 28 μm or less.• Shoe cover/roll handling capacity should be 1000 pcs or more.• Operation 220V, 50Hz.• One year warranty with parts.	<p>05 Nos. Machins</p> <p>Roll/Shoe Cover for 360000 Nos.</p>



Medical Superintendent
Faisalabad Institute of Cardiology
Faisalabad



**NUCLEAR MEDICINE ACCESSORIES
FAISALABAD INSTITUTE OF CARDIOLOGY, FAISALABAD.**

Sr. No.	Name of item	Specification	Qty
1.	Personal Electronic Dosimeter	Radiation detection gamma and X-Ray	02
2.	Patient arm Support	Compatible for Cardio MD III Gamma Camera	01
3.	Patient arm support Strap	Compatible for Cardio MD III Gamma Camera	01
4.	Patient leg rest / support	Compatible for Cardio MD III Gamma Camera	01
5.	Cardio MD table pad	Compatible for Cardio MD III Gamma Camera	01
6.	Co-57 Rectangular Flood Source	Co-57 Radioactivity: 10 mCi For Cardiac dedicated Cardio MD III Gamma Camera (Philips)	01
7.	Vile/syringe dipper	For Atomlab 400 dose calibrator	01
8.	Radiacwash	Radioactive Material decontamination liquid	01



Medical Superintendent
 Faisalabad Institute of Cardiology,
 Faisalabad.



SPECIFICATIONS OF PHOTO COPIER MACHINE
FAISALABAD INSTITUTE OF CARDIOLOGY, FAISALABAD.

Type:	Digital Copier / Network Printer, Color Scanner & Internet Fax
Gradation:	256 Levels
Front Panel	9" Full Color Capacitive Touch Pad
Copy / Print Speed	55 Pages/min (A4)
DSDF	Dual Scan Document Feeder (Std)
ADU	Automatic Duplex Unit (Std)
Resolution	1,800 (equivalent) x 600 dpi or better
Paper Size	SRA3, A5 to A3, B6 lengthwise, (12 x 48) Banner Print
Paper Capacity	4 Trays with Bypass Tray
Memory:	4096 MB (4 GB)
HDD	250GB
Zoom	25% to 400% with 1% Increments
Multiple copy	1-9999 countdown interrupting mode,
Processor	1.66MHz (dual core)
Printing Support	USB / Network
Warm up time	28 sec or less,
1st copy	3.5 sec,
USB Printing Format	TIFF, JPEG, PDF, PPTX, XLSX, etc
Paper Weight:	52 to 300 gms (By pass Tray), 52 to 256 gms (Tray 1/2)

Note: Above mentioned numeric values are approximate, furthermore equivalent parameters with different nomenclature are also acceptable.



Medical Superintendent
 Faisalabad Institute of Cardiology,
 Faisalabad.

am

9

COOLING TOWER
FAISALABAD INSTITUTE OF CARDIOLOGY, FAISALABAD.

Sr. No.	Description	Quantity
1	Cooling Tower with complete structure (1250 Ton)	1

Cooling Tower

Cooling tower should constructed as following

- | | |
|-----------------------------|--|
| (1) Capacity | 1250 TR or more |
| (2) Dimensions L x W x H | 13850 x 3940 x 3540 or more |
| (3) Inside structure | HDGI or Equivalent |
| (4) Reducer/motor support | HDGI or Equivalent |
| (5) Fan guard | HDGI or Equivalent |
| (6) Internal walkway | HDGI or Equivalent |
| (7) Casing | FRP 04 mm thick or more |
| (8) Cold water basin | FRP + HDGI |
| (9) Outlet connections | FRP |
| (10) Ladder | HDGI |
| (11) Filling | PVC 350 Microns (Size 30" x 37" or more) |
| (12) Louvers | FRP |
| (13) Bolts, Nut and washers | SS |
| (14) Access door | FRP (with SS Hinges and lock) |
| (15) Float valve | PE ball with SS rod and PVC fitting |

Fan Unit


- | | |
|-----------------|------------------------------|
| (1) No. of unit | 05 |
| (2) Type | Axial flow, Adjustable pitch |
| (3) Diameter | 2000 mm or more |
| (4) Fan speed | 300 RPM or more |
| (5) Air volume | 1550 M3/Min/Cell or more |

Fan Motor


- | | |
|---------------------|--------------------------------|
| (1) No. of unit | 05 |
| (2) Induction motor | 05 x 4.0 KW (1400 RPM) or more |
| (3) Power source | 3 Phase, 380 - 415 V, 50 Hz |
| (4) Protection | IP 55 at least |

Note: Successful bidder will supply and install the unit on site. Furthermore any accessories/parts required for installation will be arranged by the contractor. No additional payment will be made.


Note: For more detail visit the site.




Medical Superintendent
Faisalabad Institute of Cardiology,
Faisalabad.



PVMS OF MEDICAL EQUIPMENT	
Clinical Specialty	Radiological Equipment
Generic Name	X-RAY FILM PROCESSOR
Clinical Purpose	It helps in automatic processing of x-ray films providing better quality images and consuming less time.
TECHNICAL SPECIFICATIONS	
<p>High throughput free Standing type Continuous roller transport Sheet film, 10x10cm-35x43cm sizes (14"x17") High through put more than 110 sheets (35 x 43cm) Forced hot air circulation method Piping installed in the DEV and FIX racks. Processing solution temperature; Controlled by the temperature control tank, with the thermistor monitoring, with the heater heating and with the wash water cooling. Drying temperature; Controlled automatically according to a fixed temperature setting.</p>	


 16/10/2018
FOUQIR NISAR
 Bio-Medical Engineer
 Faisalabad Institute of Cardiology
 Faisalabad.


Medical Superintendent
Faisalabad Institute of Cardiology
Faisalabad.