

**MINUTES OF GRIEVANCE COMMITTEE MEETING
FAISALABAD INSTITUTE OF CARDIOLOGY, FAISALABAD**

A meeting of Grievance Committee for product of Lab Kit & Chemicals F.Y.2024-25 was held on 27-06-2024 at 11:00 am in the Office of the Medical Superintendent, Faisalabad Institute of Cardiology, Faisalabad.

Lab Kits & Chemicals

1. M/s Three Star Medical Supplies Bid

Sr	Company Name	TAC Observations	GC Observation	Remarks/Decision
1	Bid Reject of M/s Three Star Medical Supplies	Invalid license to import attached.	Renewal of license to import attached.	Accepted

2. M/s Mian Scientific Bid

Sr	Company Name	TAC Observations	GC Observation	Remarks/Decision
1	Bid Reject of M/s Mian Scientific	List of technical staff detail not attached	List of technical staff detail attached	Accepted

3. M/s Reaction Scientific (Pvt) Ltd Bid

Sr	Company Name	TAC Observations	GC Observation	Remarks/Decision
1	Bid Reject of M/s Reaction Scientific (Pvt) Ltd	Invalid drug sale license attached Punjab teaching Govt. experience not attached	Relevant documents provided	Accepted

Item #	Item Name	Brand Name	Manufactured by	TAC Observations	GC Observation	Remarks/Decision
42	Five part differential Hematology analyzer on placement with With Controls, Calibrators, Consumables and WEIRD EXTERNAL QUALITY ASSURANCE	Mindray China	Mindray China	Sufficient sample not provided for evaluation	Samples re-evaluated by GC and End user satisfied	Accepted

or

PROGRAM, ISO 17043 ACCREDITED (RIQAS, EQAS OR EQUIVALENT) AND INTERLAB COMPARISON (NEQAS, NEQAPP OR EQUIVALENT)(Tips, Cuvettes, Wash/Cleaning solutions & maintenance), INTERFACE WITH LIS / HMS IS NECESSARY.					
--	--	--	--	--	--

4. M/s Noor International

Item #	Item Name	Brand Name	Manufactured by	TAC Observations	GC Observation	Remarks/Decision
38	ABGs test with Electrolyte etc (without Lactate) INTERFACE WITH LIS / HMS IS NECESSARY.	Radiometer ABL 800 Denmark	Radiometer ABL 800 Denmark	Evaluation unit provided but there is no repeatability/precision accuracy with the existing system	End user reservation	Rejected

5. M/s Hasnain Scientific

Item #	Item Name	Brand Name	Manufactured by	TAC Observations	GC Observation	Remarks/Decision
43	PT/INR reagent for fully Automated Coagulation Analyzer With Controls, Calibrators, Consumables (Tips, Cuvettes, Wash/Cleaning solutions & maintenance) (REAGENT RENTAL BASIS OPEN TENDER) INTERFACE WITH LIS / HMS S	BehnkElektronik Germany	BehnkElektronik Germany	Sufficient sample not provided for evaluation	Samples re-evaluated by GC and End user not satisfied	Rejected

or




44	NECESSARY. APTT reagent for fully Automated Coagulation Analyzer With Controls, Calibrators, Consumables (Tips, Cuvettes, Wash/Cleaning solutions & maintenance). (REAGENT RENTAL BASIS OPEN TENDER) INTERFACE WITH LIS / HMS IS NECESSARY.	BehnkElektronik Germany	BehnkElektronik Germany	Sufficient sample not provided for evaluation	Samples re-evaluated by GC and End user not satisfied	Rejected
45	D. DIMER reagent for fully Automated Coagulation Analyzer With Controls, Calibrators, Consumables (Tips, Cuvettes, Wash/Cleaning solutions & maintenance). (REAGENT RENTAL BASIS OPEN TENDER) INTERFACE WITH LIS / HMS IS NECESSARY.	BehnkElektronik Germany	BehnkElektronik Germany	Sufficient sample not provided for evaluation	Samples re-evaluated by GC and End user not satisfied	Rejected



Mrs. Iqra Munir,
Pharmacist,
FIC, Faisalabad.
(Member)



Engr. Sharaz Ali
Bio Medical Engineer
FIC, Faisalabad.
(Member)



Dr. Rehan Riaz
Assistant Prof. of Cardiology,
FIC, Faisalabad
(Member)



Dr. Shahbaz Ahmad Khiliji
Prof. of Cardiac Surgery,
FIC, Faisalabad
(Member)



Dr. Muhammad Nadeem Akhtar
Medical Superintendent,
FIC, Faisalabad
(Member)