

**MINUTES OF THE TECHNICAL EVALUATION COMMITTEE MEETING HELD ON 22/09/2020 AT 10 AM IN THE
CONFERENCE ROOM OF F.I.C FAISALABAD FOR THE PRODUCT EVALUATION OF
CSSD Disposable Items as per list**

FOLLOWING ATTENDED

- | | |
|--------------------------|--------|
| 03. Head of Cardiology, | Member |
| 04. Dr.Husnain Zafar. | Member |
| 05. Dr Muhammad Sajid. | Member |
| 06. Miss. Iqra Liaqat, | Member |
| 07. Miss Faiza | Member |
| 08. Mr. Najam Ul Hassan, | Member |

The committee was intimated that tenders were invited through DGPR.Tender notice appeared in Daily Express, The Nation dated 04/06/2020 and on PPR.A website dated 06/06/2020 for the purchase of CSSD Disposable Items as per list.

Sr.#	Award#	Name of Item	Brand/Manufacturer	Contractor/Firm
1	1	Green Sterilization Wrapping paper (60 gms) 250 sheets/box (75x75) cm	T.N.A *Hard after processing E-Line Italy	M/s Cardiac Care ✓
			T.A Arvand E-line Srl	M/s Total Technologies Private Limited ✓
			T.A Sterisheet Stermed	M/s Summit Inter Trade ✓
			T.A Steriking Winak Finland	M/s Husnain Pharma ✓
2	2	Green Sterilization Wrapping paper (60 gms) 250 sheets/box (50x50) cm	T.N.A *Hard after processing E-Line Italy	M/s Cardiac Care ✓
			T.A Arvand E-line Srl	M/s Total Technologies Private Limited ✓
			T.A Sterisheet Stermed	M/s Summit Inter Trade ✓
3	3	Green Sterilization Wrapping paper (60 gms) 250 sheets/box (30x30) cm	T.N.A *Hard after processing E-Line Italy	M/s Cardiac Care ✓

	Name of Item	Brand/Manufacturer	Contractor/Firm
3	Green Sterilization Wrapping paper (60 gms) 250 sheets box 100x50 cm	T.A GAMS Sterilized	M/s. Summit Inter Trade
4	Flat Reel 20 cm x 200m (Packing Material)	T.A E-Line Italy T.A Avant Fire Srl L.N.A *Sample not provided Sterilization Flat Reel Skis USA L.N.A *Hard after processing VP Medical Germany T.A Sterling Wipak	M/s. Cardine Care M/s. Total Technologies Private Limited M/s. Global Marketing Services M/s. Summit Inter Trade M/s. Husnain Pharma
5	Flat Reel 20 cm x 200m (Packing Material)	T.A E-Line Italy T.A Avant Fire Srl L.N.A *Hard after processing VP Medical Germany T.A Sterling Wipak	M/s. Cardine Care M/s. Total Technologies Private Limited M/s. Summit Inter Trade M/s. Husnain Pharma
6	Flat Reel 10 cm x 200m (Packing Material)	T.A E-Line Italy T.A Avant Fire Srl L.N.A *Sample not provided Sterilization Flat Reel Skis USA	M/s. Cardine Care M/s. Total Technologies Private Limited M/s. Global Marketing Services

Handwritten signatures and initials at the bottom of the page.

#	Name of Item	Brand/Manufacturer	Contractor/Firm
6	Flat Reel 10cm x 200m (Packing Material)	T.N.A *Hard after processing VP Medical Germany	M/s. Summit Inter Trade ✓
		T.A Sterking Wipak	M/s. Husnain Pharma ✓
7	Flat Reel 15cm x 200m (Packing Material)	T.A E-Line Italy	M/s. Cardiac Care ✓
		T.A Arvand E-Line Srl	M/s. Total Technologies Private Limited ✓
		T.N.A *Sample not provided Sterilization Flat Reel Steris, USA	M/s. Global Marketing Services ✓
		T.N.A *Hard after processing VP Medical Germany	M/s. Summit Inter Trade ✓
		T.A Sterking Wipak	M/s. Husnain Pharma ✓
		T.N.A *Sample not provided E-Line Italy	M/s. Cardiac Care ✓
8	Fine Brushes for Manual Washing	T.A Batrik Surgimed	M/s. Total Technologies Private Limited ✓
		T.A E-Line Italy	M/s. Cardiac Care ✓
9	Gusseted Reel 15cm x 100m (Packing Material)	T.A E-Line Italy	M/s. Cardiac Care ✓
		T.A E-Line Italy	M/s. Total Technologies Private Limited ✓
		T.N.A *Sample not provided Sterilization Roll Steris, USA	M/s. Global Marketing Services ✓
		T.N.A *Hard after processing VP Medical Germany	M/s. Summit Inter Trade ✓


Id#	Name of Item	Brand/Manufacturer	Contractor/Firm
9	Gusseted Reel 15cm x 100m (Packing Material)	T.A Steriking Wipak	M/s: Husnain Pharma ✓
10	Gusseted Reel 40cm x 100m (Packing Material)	T.N.A *Sample not provided L-Line Italy	M/s. Cardiac Care ✓
		T.A Steriking Wipak	M/s: Husnain Pharma ✓
11	LFO Cartridge Gas (box of 12)	T.N.A *Authority letter not attached 3M-USA Germany	M/s. Total Technologies Private Limited ✓
		T.A LFO Gas Cartridge 3M USA	M/s. Anax Associates Pvt. Ltd. ✓
12	EO batch monitoring Strips (Strips for PCD) (pack of 250)	T.A Cike Germany	M/s. Total Technologies Private Limited ✓
		T.N.A *Sample not provided *Authority letter not attached Car # KH2 x 15E1/P Argentina	M/s. Summit Inter Trade ✓
13	Alkaline Chemical for washer disinfectant (5liter)	T.A Borer- Chemie Switzerland	M/s. Total Technologies Private Limited ✓
		T.N.A *Sample not provided Prolystrea 2X Concentrate Steris, USA	M/s Global Marketing Services ✓
		T.A Thermosheid Basic Dr. Schumacher GmbH, Germany	M/s. Bestech Enterprises ✓
		T.N.A *Not enough sample Acton A Acto Pharma, Turkey	M/s. Mezan International ✓
		T.N.A *Impact not good *No shining after process Labomat MA Liquid Laboratorium Dr. Deppe, Germany	M/s APSTA INT ✓

9 *[Signature]*
[Signature]
[Signature]
[Signature]

id#	Name of Item	Brand/Manufacturer	Contractor/Firm
14	Acid for washer disinfectant (5 liter)	T.A Borel-Chemie Switzerland	M/s. Total Technologies Private Limited
		T.N.A *Sample not provided Provostan 2X Enzyme Enzymatic I.MEA Steris USA	M/s. Global Marketing Services
		T.A Thermosheid-6 Dr. Seimacher GmbH, Germany	M/s. Besteca Enterprises
		T.N.A *Not compatible with machine because Alkaline not attached Genting Infection Control UK	M/s. Mediland Pakistan Pvt. Ltd
		T.N.A *No past experience with this product Actolon SN Acto Pharma	M/s. Mezan International
		T.N.A *Poor quality Laborat KS Laboratorium Dr. Deppe, Germany	M/s. APSTAINI
15	Alkalant for washer disinfectant (5 liter)	T.A Borel-Chemie Switzerland	M/s. Total Technologies Private Limited
		T.N.A *Sample not provided Large Free Pure - Pacific Steris - SA	M/s. Global Marketing Services
		T.N.A *No past experience with this product Genting Infection Control UK	M/s. Mediland Pakistan Pvt. Ltd
		T.N.A *No past experience with this product Actolon M Acto Pharma	M/s. Mezan International
		T.N.A *Not compatible Instru Care 100MI Laboratorium Dr. Deppe, Germany	M/s. APSTAINI
		T.N.A *Sample not provided SPS Medical USA	M/s. Cardiac Care
16	Label Printing Gun	T.A Gise CCT/TA	M/s. Total Technologies Private Limited

[Handwritten signatures and initials]

rd#	Name of Item	Brand/Manufacturer	Contractor/Firm
		T.N.A *Sample not provided *Authority letter not attached Cat # CG3 Argentina	M/s. Summit Inter Trade ✓
17	High level instrument disinfectant solution (5 liter)	T.N.A *Bad smell Borer- Chemie Switzerland	M/s. Total Technologies Private Limited ✓
		T.N.A *Sample not provided Liqui-jet 2 Steris, USA	M/s. Global Marketing Services ✓
		T.A Ultrademit-AF Esteer Pharma, Germany	M/s. Bestech Enterprises ✓
		T.N.A *Sample not provided ReproDIS HLD4I Medimark Scientific, UK	M/s. Husnain Pharma ✓
		T.N.A *Very bad smell Actosed PA Acto Pharma	M/s. Mezan International. ✓
		T.A Endomat Plus Viruguard Laboratorium Dr. Deppe, Germany	M/s APSTA INT ✓
18	Hand Sanitizer Gel Form	T.N.A *Sample not provided Hi Clean Alcohol Local Pakistan	M/s. Husnain Pharma ✓
19	Detergent for ultrasonic cleaner (5 liter)	T.A Borer- Chemie Switzerland	M/s. Total Technologies Private Limited ✓
		T.N.A *Sample not provided Polystica 2X Concentrate Steris, USA	M/s. Global Marketing Services ✓
		T.N.A *Discolour during process Plurazym Dr. Schumacher GmbH, Germany	M/s. Bestech Enterprises ✓
		T.A Actoclean Flusstg Acto Pharma	M/s. Mezan International. ✓



 ✓ CM Q P

ord#	Name of Item	Brand/Manufacturer	Contractor/Firm
19	Detergent for ultrasonic cleaner (5 liter)	T.A Instru ZYM Laboratorium Dr. Deppe, Germany	M/s APSTA INT ✓
20	Rust Remover (5 liter)	T.A Burer-Chemie Switzerland	M/s. Total Technologies Private Limited ✓
		T.N.A *Sample not provided Prolystion 2X Restore Descaler & Neutralizing Detergent	M/s. Global Marketing Services ✓
		T.A Actosteel Acto Pharma	M/s. Mezan International ✓
21	Steam chemical indicator (class 6)	T.N.A *Sample not according to specification (Class 5) E-Line Italy	M/s. Cardiac Care ✓
		T.N.A *Sample not provided EST 13-44 Mer Self Adhesive Steris, USA	M/s. Global Marketing Services ✓
		T.N.A *Sample not according to specification (class 4) *Authority letter not attached Cat # 1127-3YS Argentina	M/s. Summit Inter Trade ✓
		T.N.A *Discolour after process Gentinge Infection Control USA	M/s. Mediland Pakistan Pvt Ltd ✓
22	Bowie Dick test strips (for PCD) (box of 100)	T.A Gke Germany	M/s. Total Technologies Private Limited ✓
		T.N.A *Sample not provided TST Single use bowie dick Steris, USA	M/s. Global Marketing Services ✓
		T.N.A *Sample not provided *Authority letter not attached Length Range from 0.5 to 1 meter Argentina	M/s. Summit Inter Trade ✓
23	Self Contained Biological Indicator for LTO (box of 100)	T.A Gke Germany	M/s. Total Technologies Private Limited ✓


[Handwritten signatures and initials]


rd#	Name of Item	Brand/Manufacturer	Contractor/Firm	
23	Self Contained Biological indicator for FTO (box of 100)	T.N.A *Sample not provided Bio Monitor Ampuls Steris, USA	M/s. Global Marketing Services	
		T.A Biological indicator Crosstex, USA	M/s. Summit Inter Trade	
24	Self Contained biological indicator for Steam (box of 100)	T.N.A *Sample not provided SPS Medical USA	M/s. Cardiac Care	
		T.A Gke Germany	M/s. Total Technologies Private Limited	
		T.N.A *Sample not provided Bio Monitor Ampuls	M/s. Global Marketing Services	
		T.A Biological indicator Crosstex, USA	M/s. Summit Inter Trade	
		T.A Genting Infection Control USA	M/s. Mediland Pakistan Pvt. Ltd	
25	25	Sterilizer Printer cartridge 130, 134 HP	N/Q N/Q	


The committee unanimously technically approved / technically not approved the brands of the above mentioned firms in each items.


Following Items may be retendered.


Sr.No	Award #	Name Of Medicine / I / E
1	18	Hand Sanitizer Gel Form
2	21	Steam chemical indicator (class 6)
3	25	Sterilizer Printer cartridge 130, 134 HP


Mr. Najam Ul Hassan
Electronic Equipment
Care Taker.


Miss Faiza
Charge Nurse (CSSD)
F.I.C. Faisalabad


Miss. Iqra Liaqat,
Pharmacist,
F.I.C. Faisalabad


Dr. Muhammad Sajid,
AMS (Purchase /Store),
F.I.C. Faisalabad


Dr. Husnain Zafar,
Assistant Prof. of Cardiac
Surgery.

Head of Cardiology,
F.I.C. Faisalabad