

**MINUTES OF THE TECHNICAL EVALUATION COMMITTEE MEETING OF MEDICAL AND LABORATORY EQUIPMENT HELD ON 06.12.2023 AT 11:00 AM IN THE CONFERENCE ROOM, F.I.C FAISALABAD**

**FOLLOWING ATTENDED**

- |    |   |          |
|----|---|----------|
| 1. | <b>Dr. Shahid Abbas</b> , Prof. of Cardiology, F.I.C. Faisalabad.           | Chairman |
| 2. | <b>A.M.S (Purchase)</b> , F.I.C. Faisalabad.                                | Member   |
| 3. | <b>D.M.S (Purchase)</b> , F.I.C. Faisalabad.                                | Member   |
| 4. | <b>Electronic Equipment Care Taker</b> , F.I.C. Faisalabad.                 | Member   |
| 5. | <b>Pharmacist</b> , F.I.C. Faisalabad.                                      | Member   |
| 6. | <b>Mr. Haseeb Aslam</b> , Bio Medical Engineer, Allied Hospital, Faisalabad | Member   |

The Committee was intimated that tender was invited through national press i-e "Daily Dunya News" 12.11.2023 and on PPRA website dated 14.11.2023 with due date of receiving tender as 30.11.2023. The tender was opened on the same day in the presence of bidders or their authorized representative. Following firms participated in the tender process with their technical offers.

Sr. No.	Item Name	Brand , Model with	Firm participated
1.	CARDIAC MONITORS	Mindray uMEC 15 China <b>Responsive</b>	M/s Friends Traders
		Edan Intruments iM-80 China No distribution from the manufacturer <b>Non Responsive</b>	M/s Bio Tech Services
		Space Lab Healthcare XPREZZON 91393 USA <b>Responsive</b>	M/s Medical Equipment & Systems
		Edan Intruments iM-80 China No distribution from the manufacturer <b>Non Responsive</b>	M/s Clinical Life
		BioLight Q7China Arrhythmia Analysis not for paeds patients. <b>Non Responsive</b>	M/s Radiant Medical Pvt. Ltd.

2.	COLOR DOPPLER	Canon Aplio me (CUS-AME00) Japan <b>Responsive</b>	M/s Medical Equipment & Systems
		Esaote, MyLab X8 exp Italy End user reservations (Abdomen scan not satisfactory) <b>Non Responsive</b>	M/s Bio Tech Services
3.	CRASH CART TROLLEY	N6 Beijing Jingdong Tech (Laoting co. Ltd.) China <b>Responsive</b>	M/s Novatek Pakistan
4.	<u>DEFIBRILLATOR</u>	Zoll Medical M2 with pacing USA No FDA 510K <b>Non Responsive</b>	M/s Medical Equipment & Systems
		Innomed Cardio-Aid 360B Hungary <b>Responsive</b>	M/s Bio Tech Services
		OSATU, S.COOP Reanibex-700 Spain <b>Responsive</b>	M/s Radiant Medical Pvt. Ltd.
5.	ECG MACHINES	Baxter ELI-230 USA <b>Responsive</b>	M/s Radiant Medical Pvt. Ltd.
		Innomed Heart Screen 112C-1 Hungary <b>Responsive</b>	M/s Bio Tech Services
		M4Medical sp M-Trace Poland <b>Responsive</b>	M/s KASBN International

Handwritten signature or initials in blue ink.

6.	ECHOCARDIOGRAPHY MACHINE	Siemens Acuson NX3 Elite Germany <b>Responsive</b>	M/s Vertex Medical
		Esaote, MyLab X75 Italy End User Reservations (Dual Screen not available, software not user friendly, Image resolution not upto the mark <b>Non Responsive</b>	M/s Bio Tech Services
		GE Healthcare Vivid T9 USA <b>Responsive</b>	M/s Eastern Medical Technology Services
7.	PLASMA STERILIZER	GETINGE STERICOOOL 1601 Sweden <b>Responsive</b>	M/s Mediland Pakistan Pvt. Ltd.
		Sterifast E160, 2SD-B Portugal Delivery period not as required <b>Non Responsive</b>	M/s Clinical Life Inc.
		MMM Group Hyper LTS 150-2 Germany <b>Responsive</b>	M/s Total Technologies Pvt. Ltd.

on h a g

8.	MEDICAL GAS SUPPLY SYSTEM	Responsive	M/s Total Technologies Pvt. Ltd.
		Responsive	M/s Multan Chemicals Ltd
		Responsive	M/s Medi Bridge
		Responsive	M/s Radiant Medical Pvt. Ltd.
		Responsive	M/s Clinical Life Inc.

The meeting ended with the vote of thanks by the chair.

**Mr. Haseeb Aslam**  
Bio Medical Engineer  
Allied Hospital, Faisalabad.

**Electric Equipment Care Taker**  
F.I.C. Faisalabad

**Pharmacist,**  
F.I.C. Faisalabad

**D.M.S (Purchase),**  
FIC, Faisalabad.

**A.M.S (Purchase)**  
F.I.C. Faisalabad

**Dr. Shahid Abbas,**  
Prof. of Cardiology  
F.I.C. Faisalabad